
Midmark M11 - v1.0.0
Total Cycles: 4379

Name: RITTER 1
Sterilizer ID: V2178198
03/13/2024 10:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 71.7 F 0.2
4:00 82.2 F 1.2
6:00 119.2 F 3.6
8:00 151.7 F 6.5
10:00 176.4 F 9.9
12:00 200.2 F 12.5
14:00 213.0 F 4.4
16:00 223.8 F 7.7
18:00 231.9 F 11.3
20:00 228.6 F 6.1
22:00 240.9 F 11.5
24:00 235.8 F 8.7
26:00 249.3 F 14.9
28:00 260.1 F 21.0
30:00 270.5 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 271.6 F 28.7
2:00 272.0 F 28.8
3:00 272.4 F 28.9
4:00 272.3 F 29.2

Min 271.5 F 28.7
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.9 F 0.0
5:00 214.3 F 0.0
10:00 220.6 F 0.0
15:00 213.0 F 0.0
20:00 230.1 F 0.0
25:00 225.3 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 31:37
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:09:59
CYCLE COMPLETE

03/13/2024 11:34 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____