
Midmark M11 - v1.0.0
Total Cycles: 5312

Name: RITTER 1
Sterilizer ID: V2178198
02/04/2025 09:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 92.1 F 0.0
4:00 112.0 F 0.1
6:00 127.7 F 2.2
8:00 153.4 F 5.1
10:00 175.3 F 8.8
12:00 194.1 F 12.9
14:00 212.5 F 4.3
16:00 224.9 F 8.0
18:00 235.6 F 12.0
20:00 232.7 F 7.8
22:00 244.8 F 13.2
24:00 242.3 F 11.2
26:00 255.1 F 17.9
28:00 266.7 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 271.7 F 28.2
2:00 271.8 F 28.3
3:00 271.8 F 28.3
4:00 271.8 F 28.3

Min 271.6 F 28.0
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
Dry Cycle Aborted

FILLING: 0:41
HEATING: 29:25
STERILIZING: 4:00
VENTING: 3:02
DRYING: 0:08
TOTAL CYCLE: 00:37:16

02/04/2025 09:53 AM
ACCEPTED:
REJECTED:
Reason: -----
Operator: -----