
Midmark M11 - v1.0.5
Total Cycles: 2546

Name: RITTER2
Sterilizer ID: V2313228
06/21/2023 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.2 F 0.0
2:00 190.8 F 4.0
4:00 204.4 F 9.3
6:00 222.0 F 7.8
8:00 232.1 F 10.8
10:00 233.4 F 8.4
12:00 240.2 F 10.5
14:00 247.1 F 14.1
16:00 259.6 F 21.0
18:00 270.8 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.7 F 29.5
2:00 271.6 F 28.7
3:00 272.2 F 29.6
4:00 272.5 F 29.3

Min 271.4 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 214.6 F 0.0
10:00 223.1 F 0.0
15:00 224.0 F 0.0
20:00 223.5 F 0.0
25:00 224.0 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 18:49
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

06/21/2023 12:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____