

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2189

Name: RITTER2  
Sterilizer ID: V2313228  
03/28/2023 04:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.2 F 0.0  
2:00 169.0 F 3.5  
4:00 185.4 F 7.6  
6:00 204.7 F 12.6  
8:00 221.2 F 7.2  
10:00 233.9 F 12.5  
12:00 236.3 F 10.1  
14:00 236.4 F 8.8  
16:00 250.4 F 15.9  
18:00 261.9 F 22.6  
20:00 272.6 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.5 F 29.5  
2:00 271.4 F 29.1  
3:00 272.4 F 29.6  
4:00 271.5 F 28.9

Min 271.3 F 28.9  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
5:00 211.2 F 0.0  
10:00 217.7 F 0.0  
15:00 214.3 F 0.0  
20:00 214.1 F 0.0  
25:00 215.2 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 20:24  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 00:59:25  
CYCLE COMPLETE

03/28/2023 05:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_