
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000846
15:09 06/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:56
STERILIZING 5:20
136.0 C 227kPa 5:20
136.2 C 232kPa 5:50
136.4 C 233kPa 6:20
136.4 C 232kPa 6:50
136.1 C 229kPa 7:20
136.4 C 233kPa 7:50
136.4 C 232kPa 8:20
136.4 C 232kPa 8:50
136.4 C 232kPa 8:50
VENTING 8:50
AIR DRYING 9:48
CYCLE COMPLETE 24:48

Digital Signature #
8C91705CDE255D77
