

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1358

Name: RITTER 1  
Sterilizer ID: V2178198  
11/11/2021 07:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.5 F 0.0  
2:00 73.0 F 0.0  
4:00 78.0 F 0.0  
6:00 115.6 F 0.1  
8:00 147.8 F 0.9  
10:00 174.4 F 3.6  
12:00 214.1 F 6.7  
14:00 217.4 F 10.7  
16:00 232.1 F 6.7  
18:00 228.3 F 8.3  
20:00 239.1 F 12.7  
22:00 238.3 F 9.4  
24:00 240.1 F 9.5  
26:00 247.3 F 13.4  
28:00 259.0 F 19.8  
30:00 269.6 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 27.8  
1:00 272.2 F 28.4  
2:00 272.0 F 28.4  
3:00 272.4 F 28.4  
4:00 272.7 F 28.8

Min 271.0 F 27.7  
Max 273.0 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
5:00 200.2 F 0.0  
10:00 202.8 F 0.0  
15:00 202.1 F 0.0  
20:00 216.5 F 0.0  
25:00 232.2 F 0.0  
30:00 230.1 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 30:59  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:09:02  
CYCLE COMPLETE

11/11/2021 09:06 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_