
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005854
8:29 28/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:47
PRESSURIZING 1:52
STERILIZING 2:46
136.0 C 220kPa 2:46
136.3 C 223kPa 3:16
136.4 C 223kPa 3:46
136.1 C 221kPa 4:16
136.4 C 223kPa 4:46
136.4 C 222kPa 5:16
136.4 C 223kPa 5:46
136.4 C 223kPa 6:16
136.4 C 223kPa 6:16
VENTING 6:16
AIR DRYING 7:09
CYCLE COMPLETE 7:25

Digital Signature #
5F104E152CD3A3D1
