
Midmark M11 - v1.0.0
Total Cycles: 2187

Name: RITTER 1
Sterilizer ID: V2178198
07/27/2022 01:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.8 F 0.0
2:00 167.2 F 3.3
4:00 178.4 F 7.6
6:00 201.4 F 12.4
8:00 214.3 F 6.7
10:00 231.3 F 11.7
12:00 233.0 F 8.1
14:00 242.3 F 11.9
16:00 244.0 F 12.2
18:00 255.8 F 18.7
20:00 266.9 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.6 F 29.4
2:00 271.9 F 28.7
3:00 271.8 F 29.1
4:00 272.5 F 29.3

Min 271.4 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 216.6 F 0.0
10:00 234.1 F 0.0
15:00 236.4 F 0.0
20:00 225.3 F 0.0
25:00 229.5 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:35
CYCLE COMPLETE

07/27/2022 02:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____