
Midmark M11 - v1.0.0
Total Cycles: 3355

Name: RITTER 1
Sterilizer ID: V2178198
06/08/2023 07:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.0 F 0.0
2:00 78.9 F 0.0
4:00 94.2 F 0.0
6:00 110.1 F 0.0
8:00 146.8 F 2.1
10:00 173.6 F 5.1
12:00 194.2 F 8.4
14:00 208.1 F 12.5
16:00 216.0 F 5.6
18:00 225.8 F 9.4
20:00 236.1 F 12.9
22:00 236.6 F 9.8
24:00 237.6 F 9.2
26:00 246.4 F 13.5
28:00 257.9 F 19.6
30:00 268.3 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.6 F 29.0
2:00 271.5 F 28.2
3:00 272.6 F 29.1
4:00 272.1 F 28.6

Min 271.3 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 208.6 F 0.0
10:00 223.4 F 0.0
15:00 222.0 F 0.0
20:00 230.1 F 0.0
25:00 221.3 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:18
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:09:22
CYCLE COMPLETE

06/08/2023 08:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____