
Midmark M11 - v1.0.0
Total Cycles: 1751

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2022 08:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.7 F 0.0
2:00 164.6 F 1.1
4:00 165.9 F 4.1
6:00 188.2 F 8.2
8:00 208.4 F 13.3
10:00 221.7 F 7.3
12:00 235.4 F 12.5
14:00 236.6 F 9.6
16:00 236.7 F 8.7
18:00 248.5 F 14.6
20:00 260.8 F 21.5
22:00 272.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.5 F 29.3
2:00 272.0 F 28.7
3:00 271.2 F 28.4
4:00 272.4 F 29.3

Min 271.1 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 199.7 F 0.0
10:00 198.8 F 0.0
15:00 200.5 F 0.0
20:00 200.7 F 0.0
25:00 228.2 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:35
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:36
CYCLE COMPLETE

03/28/2022 09:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____