
Midmark M11 - v1.0.5
Total Cycles: 2808

Name: RITTER2
Sterilizer ID: V2313228
08/29/2023 09:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.4 F 0.0
2:00 72.0 F 0.0
4:00 91.0 F 0.0
6:00 121.5 F 0.0
8:00 151.7 F 0.0
10:00 179.4 F 0.1
12:00 191.6 F 2.3
14:00 204.0 F 4.8
16:00 214.8 F 7.4
18:00 228.0 F 10.5
20:00 233.1 F 9.1
22:00 234.9 F 9.9
24:00 240.9 F 10.4
26:00 243.5 F 12.6
28:00 241.8 F 11.3
30:00 253.3 F 17.2
32:00 263.4 F 23.2
34:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.3 F 28.9
2:00 272.5 F 29.4
3:00 271.5 F 28.5
4:00 272.5 F 29.0

Min 271.5 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 203.9 F 0.0
10:00 204.6 F 0.0
15:00 204.7 F 0.0
20:00 204.5 F 0.0
25:00 203.9 F 0.0
30:00 201.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 34:19
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:12:53
CYCLE COMPLETE

08/29/2023 10:21 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____