

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2944

Name: RITTER 1  
Sterilizer ID: V2178198  
02/27/2023 09:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.4 F 0.0  
2:00 192.8 F 2.4  
4:00 211.0 F 6.0  
6:00 220.8 F 10.9  
8:00 222.7 F 5.6  
10:00 236.9 F 11.2  
12:00 232.7 F 8.0  
14:00 243.6 F 11.8  
16:00 244.9 F 12.6  
18:00 257.3 F 19.1  
20:00 268.5 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.4 F 29.1  
2:00 271.8 F 28.5  
3:00 271.4 F 28.4  
4:00 272.5 F 29.1  
  
Min 271.1 F 28.1  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.7 F 0.0  
5:00 217.1 F 0.0  
10:00 243.3 F 0.0  
15:00 230.7 F 0.0  
20:00 226.0 F 0.0  
25:00 225.0 F 0.0  
30:00 223.1 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 21:17  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:37  
CYCLE COMPLETE

02/27/2023 10:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_