
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001577
14:57 11/11/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:13
PRESSURIZING 6:11
STERILIZING 7:42
136.0 C 220kPa 7:42
136.4 C 220kPa 8:12
136.4 C 223kPa 8:42
136.2 C 218kPa 9:12
136.4 C 222kPa 9:43
136.3 C 222kPa 10:13
136.1 C 220kPa 10:43
136.4 C 222kPa 11:12
VENTING 11:12
AIR DRYING 12:23
CYCLE COMPLETE 42:23

Digital Signature #
609E6C6C477EAE79
