
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002506
15:09 10/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:00
STERILIZING 3:02
136.4 C 225kPa 3:02
136.2 C 224kPa 3:32
136.2 C 225kPa 4:02
136.4 C 226kPa 4:32
136.4 C 226kPa 5:02
136.3 C 225kPa 5:32
136.4 C 227kPa 6:02
136.4 C 226kPa 6:32
136.4 C 226kPa 6:32
VENTING 6:32
AIR DRYING 7:27
CYCLE COMPLETE 27:59

Digital Signature #
33DED9E8C40CB636
