
Midmark M11 - v1.0.5
Total Cycles: 1554

Name: RITTER2
Sterilizer ID: V2313228
10/03/2022 01:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.2 F 0.0
2:00 199.0 F 1.0
4:00 208.1 F 6.3
6:00 228.8 F 11.8
8:00 232.6 F 9.0
10:00 239.3 F 9.6
12:00 243.1 F 11.8
14:00 246.8 F 13.9
16:00 259.5 F 21.0
18:00 271.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.7 F 29.4
2:00 271.6 F 28.8
3:00 272.7 F 29.6
4:00 271.9 F 28.9

Min 271.5 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 230.0 F 0.0
10:00 239.0 F 0.0
15:00 240.5 F 0.0
20:00 226.4 F 0.0
25:00 223.0 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 18:42
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:57:50
CYCLE COMPLETE

10/03/2022 02:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____