

-----  
Midmark M11 - v1.0.5  
Total Cycles: 739

Name: RITTER2  
Sterilizer ID: V2313228  
01/31/2022 04:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.9 F 0.0  
2:00 120.9 F 0.1  
4:00 149.0 F 2.3  
6:00 180.4 F 6.5  
8:00 203.8 F 11.2  
10:00 216.5 F 6.1  
12:00 231.6 F 11.5  
14:00 233.2 F 8.9  
16:00 239.0 F 10.2  
18:00 247.0 F 14.3  
20:00 259.7 F 21.1  
22:00 271.0 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.0 F 29.1  
2:00 272.5 F 29.6  
3:00 271.6 F 29.1  
4:00 272.4 F 29.3

Min 271.4 F 28.8  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.5 F 0.0  
5:00 220.1 F 0.0  
10:00 224.9 F 0.0  
15:00 220.9 F 0.0  
20:00 222.2 F 0.0  
25:00 221.0 F 0.0  
30:00 220.8 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 22:39  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:00:49  
CYCLE COMPLETE

01/31/2022 05:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_