

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1258

Name: RITTER 1  
Sterilizer ID: V2178198  
10/13/2021 06:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.7 F 0.0  
2:00 75.7 F 0.0  
4:00 79.8 F 0.0  
6:00 122.0 F 0.0  
8:00 149.5 F 0.1  
10:00 182.3 F 0.2  
12:00 201.3 F 0.3  
14:00 213.8 F 2.9  
16:00 221.5 F 5.3  
18:00 229.2 F 8.0  
20:00 237.0 F 11.4  
22:00 232.0 F 7.1  
24:00 243.9 F 12.6  
26:00 240.9 F 10.7  
28:00 237.5 F 9.1  
30:00 251.3 F 15.6  
32:00 262.2 F 21.8  
34:00 272.3 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 272.6 F 28.6  
2:00 271.5 F 27.8  
3:00 272.4 F 28.6  
4:00 272.2 F 28.3

Min 271.4 F 27.7  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.3 F 0.0  
5:00 202.0 F 0.0  
10:00 212.5 F 0.0  
15:00 207.9 F 0.0  
20:00 203.4 F 0.0  
25:00 212.7 F 0.0  
30:00 226.6 F 0.0

Drying Complete

FILLING: 0:54  
HEATING: 34:31  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:12:45  
CYCLE COMPLETE

10/13/2021 07:57 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01