
Midmark M11 - v1.0.0
Total Cycles: 5360

Name: RITTER 1
Sterilizer ID: V2178198
02/24/2025 02:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.8 F 0.0
2:00 130.8 F 1.6
4:00 139.3 F 4.5
6:00 165.2 F 7.7
8:00 186.1 F 11.3
10:00 201.4 F 2.5
12:00 213.7 F 5.8
14:00 225.5 F 9.6
16:00 237.3 F 11.9
18:00 236.1 F 9.9
20:00 238.2 F 9.9
22:00 244.9 F 13.1
24:00 255.9 F 18.8
26:00 265.8 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.2 F 29.4
2:00 272.6 F 29.3
3:00 272.2 F 29.0
4:00 271.6 F 28.6

Min 271.4 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 192.3 F 0.0
10:00 187.4 F 0.0
15:00 186.5 F 0.0
20:00 189.3 F 0.0
25:00 193.0 F 0.0
30:00 195.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 27:57
STERILIZING: 4:00
VENTING: 4:05
DRYING: 30:00
TOTAL CYCLE: 01:06:55
CYCLE COMPLETE

02/24/2025 03:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____