
Midmark M11 - v1.0.0
Total Cycles: 1810

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2022 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.6 F 0.0
2:00 194.0 F 4.1
4:00 197.8 F 9.1
6:00 213.3 F 8.0
8:00 227.0 F 10.0
10:00 235.1 F 7.8
12:00 243.8 F 13.1
14:00 243.2 F 11.6
16:00 256.1 F 18.3
18:00 267.3 F 25.4
20:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.8 F 29.4
2:00 271.9 F 28.7
3:00 272.1 F 29.3
4:00 272.6 F 29.3

Min 271.5 F 28.6
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 231.6 F 0.0
10:00 223.0 F 0.0
15:00 221.5 F 0.0
20:00 221.5 F 0.0
25:00 224.8 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 20:08
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:60
CYCLE COMPLETE

04/11/2022 01:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____