
Midmark M11 - v1.0.5
Total Cycles: 1454

Name: RITTER2
Sterilizer ID: V2313228
09/06/2022 06:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.6 F 0.0
2:00 83.7 F 0.0
4:00 108.7 F 0.0
6:00 124.9 F 0.0
8:00 140.5 F 1.6
10:00 178.9 F 3.6
12:00 211.0 F 5.9
14:00 229.2 F 9.2
16:00 240.4 F 13.1
18:00 233.8 F 8.0
20:00 244.4 F 12.7
22:00 239.2 F 10.0
24:00 237.3 F 8.9
26:00 249.2 F 14.8
28:00 260.4 F 21.1
30:00 270.5 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.6 F 29.0
2:00 271.1 F 28.5
3:00 272.6 F 29.0
4:00 271.0 F 28.4

Min 271.0 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 157.8 F 0.0
10:00 151.5 F 0.0
15:00 143.3 F 0.0
20:00 139.1 F 0.0
25:00 135.6 F 0.0
30:00 132.6 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 30:49
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:08:51
CYCLE COMPLETE

09/06/2022 08:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____