
Midmark M11 - v1.0.5
Total Cycles: 2542

Name: RITTER2
Sterilizer ID: V2313228
06/21/2023 07:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 77.5 F 0.0
4:00 87.9 F 0.0
6:00 100.1 F 0.1
8:00 136.8 F 2.8
10:00 164.2 F 6.2
12:00 194.0 F 9.9
14:00 214.1 F 10.4
16:00 219.3 F 7.1
18:00 229.5 F 11.1
20:00 227.7 F 7.0
22:00 239.6 F 12.0
24:00 237.3 F 9.6
26:00 248.9 F 15.4
28:00 259.4 F 21.2
30:00 269.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.3 F 29.7
2:00 271.1 F 29.2
3:00 272.3 F 29.8
4:00 272.2 F 29.8

Min 271.0 F 29.1
Max 273.2 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 203.8 F 0.0
10:00 216.0 F 0.0
15:00 220.0 F 0.0
20:00 218.1 F 0.0
25:00 220.0 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:09
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:07
CYCLE COMPLETE

06/21/2023 08:29 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____