
Midmark M11 - v1.0.0
Total Cycles: 4855

Name: RITTER 1
Sterilizer ID: V2178198
08/27/2024 01:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 80.2 F 0.0
2:00 81.0 F 0.0
4:00 96.7 F 0.3
6:00 132.9 F 1.4
8:00 163.2 F 1.1
10:00 192.1 F 4.5
12:00 204.0 F 8.5
14:00 217.4 F 13.0
16:00 225.7 F 6.9
18:00 236.8 F 11.6
20:00 234.7 F 8.5
22:00 240.7 F 10.3
24:00 246.2 F 13.2
26:00 258.6 F 19.9
28:00 269.7 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 271.7 F 28.1
2:00 272.1 F 28.4
3:00 272.6 F 28.7
4:00 272.3 F 28.9

Min 271.4 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.5 F 0.0
5:00 231.6 F 0.0
10:00 219.8 F 0.0
15:00 176.6 F 0.0
20:00 175.3 F 0.0
25:00 174.3 F 0.0
30:00 151.2 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 28:51
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:06:41
CYCLE COMPLETE

08/27/2024 02:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____