
Midmark M11 - v1.0.0
Total Cycles: 3853

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2023 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 94.0 F 0.0
4:00 90.5 F 0.0
6:00 113.8 F 0.0
8:00 165.4 F 1.2
10:00 193.2 F 4.2
12:00 208.0 F 8.1
14:00 222.1 F 12.1
16:00 226.4 F 6.0
18:00 234.7 F 10.0
20:00 239.8 F 9.5
22:00 243.0 F 11.8
24:00 240.0 F 10.0
26:00 253.8 F 16.9
28:00 265.1 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.0 F 28.4
2:00 272.7 F 28.5
3:00 272.4 F 28.3
4:00 272.1 F 28.1

Min 271.5 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 234.4 F 0.0
Dry Cycle Aborted

FILLING: 0:53
HEATING: 29:37
STERILIZING: 4:00
VENTING: 2:54
DRYING: 1:03
TOTAL CYCLE: 00:38:27

10/03/2023 10:17 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____