
Midmark M11 - v1.0.0
Total Cycles: 5046

Name: RITTER 1
Sterilizer ID: V2178198
11/01/2024 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.5 F 0.0
2:00 130.5 F 0.3
4:00 151.1 F 0.6
6:00 182.2 F 3.1
8:00 207.8 F 6.9
10:00 213.1 F 11.3
12:00 221.8 F 5.6
14:00 233.0 F 10.7
16:00 232.2 F 7.2
18:00 244.9 F 13.5
20:00 243.6 F 12.0
22:00 256.0 F 18.5
24:00 267.0 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 271.7 F 28.8
2:00 272.2 F 28.8
3:00 271.2 F 28.2
4:00 272.0 F 28.9

Min 271.0 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 225.1 F 0.0
10:00 217.9 F 0.0
15:00 221.7 F 0.0
20:00 222.3 F 0.0
25:00 214.2 F 0.0
30:00 231.1 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 25:23
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:03:48
CYCLE COMPLETE

11/01/2024 02:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____