
Midmark M11 - v1.0.0
Total Cycles: 4599

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2024 01:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 70.3 F 0.3
4:00 76.8 F 1.1
6:00 102.1 F 3.2
8:00 131.9 F 5.8
10:00 160.3 F 9.1
12:00 180.9 F 12.6
14:00 206.6 F 3.2
16:00 212.9 F 6.4
18:00 224.1 F 9.6
20:00 232.3 F 13.4
22:00 234.0 F 8.9
24:00 244.9 F 12.3
26:00 242.1 F 11.3
28:00 253.6 F 17.1
30:00 263.8 F 23.1
32:00 272.2 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.1 F 28.3
2:00 272.4 F 28.6
3:00 272.6 F 28.9
4:00 272.1 F 28.8

Min 271.3 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 199.7 F 0.0
10:00 196.9 F 0.0
15:00 180.8 F 0.0
20:00 194.7 F 0.0
25:00 186.4 F 0.0
30:00 176.3 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:09
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:10:17
CYCLE COMPLETE

05/31/2024 02:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____