
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006623
12:21 15/11/2023
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 4:10
STERILIZING 5:04
136.4 C 220kPa 5:04
136.5 C 225kPa 5:34
136.1 C 222kPa 6:04
136.4 C 221kPa 6:34
136.4 C 222kPa 7:04
136.4 C 222kPa 7:34
136.4 C 225kPa 8:04
136.4 C 222kPa 8:34
136.4 C 217kPa 9:04
136.4 C 221kPa 9:34
136.4 C 223kPa 10:04
136.3 C 223kPa 10:34
136.4 C 223kPa 11:04
136.1 C 220kPa 11:34
136.4 C 223kPa 12:04
136.4 C 221kPa 12:34
136.0 C 222kPa 13:04
136.1 C 218kPa 13:34
136.3 C 222kPa 14:04
136.2 C 223kPa 14:34
136.3 C 224kPa 15:04
136.3 C 224kPa 15:04
VENTING 15:04
AIR DRYING 15:55
CYCLE COMPLETE 16:12

Digital Signature #

545C8B448EFC7BA0
