

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3260

Name: RITTER 1  
Sterilizer ID: V2178198  
05/19/2023 07:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 79.7 F 0.0  
2:00 87.6 F 0.0  
4:00 98.9 F 0.0  
6:00 124.5 F 0.0  
8:00 153.5 F 0.8  
10:00 174.6 F 3.9  
12:00 189.9 F 7.1  
14:00 205.9 F 10.3  
16:00 218.4 F 9.6  
18:00 223.3 F 7.0  
20:00 233.4 F 10.9  
22:00 233.2 F 7.6  
24:00 243.1 F 12.2  
26:00 239.7 F 10.0  
28:00 251.3 F 15.9  
30:00 261.7 F 21.8  
32:00 271.4 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.5 F 28.9  
2:00 271.6 F 28.1  
3:00 272.2 F 29.0  
4:00 272.2 F 28.6  
  
Min 271.3 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.2 F 0.0  
5:00 224.9 F 0.0  
10:00 226.3 F 0.0  
15:00 224.7 F 0.0  
20:00 229.6 F 0.0  
25:00 231.0 F 0.0  
30:00 229.8 F 0.0  
  
Drying Complete

FILLING: 0:53  
HEATING: 32:42  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:10:52  
CYCLE COMPLETE

05/19/2023 08:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_