
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004730
10:18 21/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 1:59
STERILIZING 2:58
136.4 C 220kPa 2:58
136.4 C 222kPa 3:28
136.4 C 224kPa 3:58
136.1 C 222kPa 4:28
136.4 C 224kPa 4:58
136.2 C 222kPa 5:28
136.3 C 222kPa 5:58
136.1 C 222kPa 6:28
136.1 C 222kPa 6:28
VENTING 6:28
AIR DRYING 8:00
CYCLE COMPLETE 17:37

Digital Signature #
089E0C89CA6BC137
