

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 006292  
11:33 12/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:28  
STERILIZING 3:26  
136.0 C 220kPa 3:26  
136.1 C 223kPa 3:56  
136.3 C 224kPa 4:26  
136.4 C 224kPa 4:56  
136.4 C 221kPa 5:26  
136.1 C 222kPa 5:56  
136.4 C 224kPa 6:26  
136.1 C 221kPa 6:56  
136.1 C 221kPa 6:56  
VENTING 6:56  
AIR DRYING 7:53  
CYCLE COMPLETE 7:58

-----  
Digital Signature #  
806BA12424E5FE8C  
-----