

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2054

Name: RITTER 1  
Sterilizer ID: V2178198  
06/24/2022 01:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 220.4 F 0.0  
2:00 216.8 F 3.8  
4:00 225.4 F 9.0  
6:00 228.3 F 7.6  
8:00 235.5 F 10.1  
10:00 235.1 F 8.0  
12:00 245.9 F 13.4  
14:00 244.1 F 12.1  
16:00 256.5 F 18.6  
18:00 267.3 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.4 F 28.9  
2:00 271.5 F 28.2  
3:00 271.8 F 28.7  
4:00 272.3 F 28.8  
  
Min 271.2 F 28.1  
Max 272.4 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.6 F 0.0  
5:00 232.1 F 0.0  
10:00 226.6 F 0.0  
15:00 229.3 F 0.0  
20:00 232.4 F 0.0  
25:00 228.4 F 0.0  
30:00 228.7 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 19:28  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:40  
CYCLE COMPLETE

06/24/2022 02:37 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_