
Midmark M11 - v1.0.0
Total Cycles: 1839

Name: RITTER 1
Sterilizer ID: V2178198
04/19/2022 09:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.0 F 0.0
2:00 189.5 F 3.2
4:00 195.8 F 7.9
6:00 215.7 F 13.4
8:00 225.0 F 8.3
10:00 237.9 F 12.3
12:00 238.8 F 11.3
14:00 238.1 F 9.8
16:00 252.1 F 16.9
18:00 264.4 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.6 F 29.6
2:00 271.9 F 28.9
3:00 271.6 F 29.1
4:00 272.7 F 29.5

Min 271.3 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 216.4 F 0.0
10:00 222.5 F 0.0
15:00 229.7 F 0.0
20:00 233.8 F 0.0
25:00 227.5 F 0.0
30:00 224.1 F 0.0

Drying Complete

FILLING: 1:38
HEATING: 19:53
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

04/19/2022 10:11 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____