
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000443
13:44 28/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:42
STERILIZING 3:40
136.4 C 220kPa 3:40
136.3 C 231kPa 4:10
136.3 C 232kPa 4:40
136.3 C 230kPa 5:10
136.2 C 232kPa 5:40
136.3 C 232kPa 6:10
136.4 C 233kPa 6:40
136.4 C 233kPa 7:10
136.4 C 233kPa 7:10
VENTING 7:10
AIR DRYING 8:02
CYCLE COMPLETE 8:17

Digital Signature #
7FC55DBD44A00CE1
