

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004127  
14:04 27/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:24  
PRESSURIZING 2:28  
STERILIZING 3:24  
136.5 C 220kPa 3:24  
136.2 C 220kPa 3:54  
136.4 C 225kPa 4:24  
136.4 C 225kPa 4:54  
136.4 C 222kPa 5:24  
136.3 C 223kPa 5:54  
136.1 C 222kPa 6:24  
136.4 C 224kPa 6:54  
136.4 C 224kPa 6:54  
VENTING 6:55  
AIR DRYING 7:49  
CYCLE COMPLETE 8:30

-----  
Digital Signature #  
DDA81D151E45E63B  
-----