
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005518
10:53 11/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:40
STERILIZING 3:46
136.5 C 221kPa 3:46
136.4 C 225kPa 4:16
136.4 C 221kPa 4:46
136.4 C 224kPa 5:16
136.4 C 223kPa 5:46
136.4 C 224kPa 6:16
136.1 C 218kPa 6:46
136.1 C 221kPa 7:16
136.1 C 221kPa 7:16
VENTING 7:16
AIR DRYING 8:06
CYCLE COMPLETE 38:06

Digital Signature #
E66B3ED92A98ACA3
