

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 005643  
15:23 31/05/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:20  
PRESSURIZING 2:25  
STERILIZING 3:16  
136.4 C 221kPa 3:16  
136.4 C 225kPa 3:46  
136.4 C 224kPa 4:16  
136.4 C 222kPa 4:46  
136.1 C 220kPa 5:16  
136.4 C 225kPa 5:46  
136.3 C 223kPa 6:16  
136.0 C 224kPa 6:46  
136.0 C 224kPa 6:46  
VENTING 6:46  
AIR DRYING 7:35  
CYCLE COMPLETE 8:00

-----  
Digital Signature #  
164F9EB803F0FAFB  
-----