

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1059

Name: RITTER 1  
Sterilizer ID: V2178198  
07/27/2021 10:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.2 F 0.0  
2:00 152.2 F 1.7  
4:00 163.4 F 4.8  
6:00 191.2 F 8.8  
8:00 219.2 F 13.5  
10:00 221.9 F 8.4  
12:00 234.7 F 13.5  
14:00 236.1 F 10.5  
16:00 234.3 F 8.2  
18:00 249.4 F 15.0  
20:00 260.9 F 21.7  
22:00 271.7 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.1 F 29.0  
2:00 272.6 F 29.3  
3:00 271.8 F 29.4  
4:00 271.6 F 28.9

Min 271.4 F 28.8  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.4 F 0.0  
5:00 233.9 F 0.0  
10:00 224.9 F 0.0  
15:00 226.9 F 0.0  
20:00 226.7 F 0.0  
25:00 224.3 F 0.0  
30:00 217.3 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 23:19  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:01:29  
CYCLE COMPLETE

07/27/2021 11:50 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01