
Midmark M11 - v1.0.0
Total Cycles: 1245

Name: RITTER 1
Sterilizer ID: V2178198
10/07/2021 04:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.2 F 0.0
2:00 180.3 F 0.1
4:00 196.7 F 1.3
6:00 213.9 F 4.6
8:00 228.3 F 8.3
10:00 240.0 F 12.4
12:00 238.0 F 9.4
14:00 240.3 F 10.0
16:00 247.1 F 13.4
18:00 244.9 F 12.3
20:00 257.2 F 18.7
22:00 268.2 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 27.8
1:00 272.5 F 28.5
2:00 272.0 F 27.8
3:00 272.4 F 28.5
4:00 272.3 F 28.1

Min 271.5 F 27.6
Max 272.9 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 226.9 F 0.0
10:00 224.7 F 0.0
15:00 224.7 F 0.0
20:00 219.3 F 0.0
25:00 219.3 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:09
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:22
CYCLE COMPLETE

10/07/2021 05:42 PM
ACCEPTED: X
REJECTED:
Operator: U02