

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002576  
10:48 14/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:43  
STERILIZING 3:54  
136.5 C 224kPa 3:54  
136.4 C 224kPa 4:24  
136.4 C 225kPa 4:54  
136.3 C 223kPa 5:24  
136.4 C 225kPa 5:54  
136.4 C 225kPa 6:24  
136.2 C 223kPa 6:54  
136.4 C 224kPa 7:24  
136.4 C 224kPa 7:24  
VENTING 7:24  
AIR DRYING 8:14  
CYCLE COMPLETE 24:37  
-----

Digital Signature #  
E4610CB38679292D  
-----