
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000831
15:46 21/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:03
STERILIZING 4:18
136.0 C 229kPa 4:18
136.3 C 233kPa 4:48
136.4 C 229kPa 5:18
136.1 C 230kPa 5:48
136.3 C 231kPa 6:18
136.4 C 233kPa 6:48
136.2 C 230kPa 7:18
136.3 C 232kPa 7:48
136.3 C 232kPa 7:48
VENTING 7:48
AIR DRYING 8:43
CYCLE COMPLETE 16:01

Digital Signature #
8D3BB04908DCEDD3
