
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 008021
15:58 04/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:47
STERILIZING 4:52
136.4 C 222kPa 4:52
136.1 C 221kPa 5:22
136.1 C 222kPa 5:52
136.4 C 224kPa 6:22
136.0 C 217kPa 6:52
136.4 C 224kPa 7:22
136.4 C 223kPa 7:52
136.4 C 224kPa 8:22
136.4 C 224kPa 8:22
VENTING 8:22
AIR DRYING 9:21
CYCLE COMPLETE 10:10

Digital Signature #
1610D92A526A95AB
