
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 000573
15:19 09/08/2021
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 5:41
STERILIZING 6:30
136.0 C 227kPa 6:30
136.2 C 229kPa 7:00
136.3 C 231kPa 7:30
136.3 C 230kPa 8:00
136.3 C 229kPa 8:30
136.3 C 231kPa 9:00
136.3 C 230kPa 9:30
136.3 C 229kPa 10:00
136.2 C 230kPa 10:30
136.3 C 231kPa 11:00
136.0 C 229kPa 11:30
136.4 C 233kPa 12:00
136.1 C 230kPa 12:30
136.4 C 233kPa 13:00
136.4 C 233kPa 13:30
136.4 C 233kPa 14:00
136.4 C 233kPa 14:30
136.4 C 230kPa 15:00
136.4 C 231kPa 15:30
136.3 C 231kPa 16:00
136.4 C 233kPa 16:30
136.4 C 233kPa 16:30
VENTING 16:30
AIR DRYING 17:24
CYCLE COMPLETE 32:24

Digital Signature #
085C9CE334181286
