
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000515
9:12 28/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:31
PRESSURIZING 3:34
STERILIZING 4:43
136.0 C 220kPa 4:43
136.4 C 224kPa 5:13
136.4 C 224kPa 5:43
136.3 C 223kPa 6:13
136.4 C 222kPa 6:43
136.2 C 224kPa 7:13
136.2 C 221kPa 7:43
136.4 C 224kPa 8:13
136.4 C 224kPa 8:13
VENTING 8:13
AIR DRYING 9:05
CYCLE COMPLETE 9:07

Digital Signature #
7A4D36CED2BF97A1
