

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2756

Name: RITTER 1  
Sterilizer ID: V2178198  
01/10/2023 09:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 210.5 F 0.0  
2:00 207.1 F 0.0  
4:00 207.4 F 3.7  
6:00 213.5 F 9.1  
8:00 227.0 F 8.6  
10:00 236.7 F 10.4  
12:00 234.6 F 7.9  
14:00 242.4 F 11.1  
16:00 247.1 F 13.6  
18:00 259.7 F 20.5  
20:00 271.0 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.5 F 28.9  
2:00 271.9 F 28.2  
3:00 271.7 F 28.4  
4:00 272.5 F 28.8  
  
Min 271.3 F 27.9  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.7 F 0.0  
5:00 230.2 F 0.0  
10:00 240.1 F 0.0  
15:00 231.2 F 0.0  
20:00 227.4 F 0.0  
25:00 229.3 F 0.0  
30:00 225.4 F 0.0  
Drying Complete

FILLING: 1:30  
HEATING: 20:46  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:28  
CYCLE COMPLETE

01/10/2023 10:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_