
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006165
13:49 18/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:00
STERILIZING 2:58
136.4 C 221kPa 2:58
135.9 C 221kPa 3:28
136.2 C 223kPa 3:58
136.4 C 224kPa 4:28
136.4 C 223kPa 4:58
136.4 C 224kPa 5:28
136.4 C 224kPa 5:58
136.4 C 225kPa 6:28
136.4 C 225kPa 6:28
VENTING 6:28
AIR DRYING 7:21
CYCLE COMPLETE 21:39

Digital Signature #
7BAF204C1981AB0D
