
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000480
14:36 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:47
PRESSURIZING 1:52
STERILIZING 2:55
136.4 C 220kPa 2:55
136.4 C 224kPa 3:25
136.2 C 223kPa 3:55
136.0 C 222kPa 4:25
136.1 C 219kPa 4:55
136.4 C 225kPa 5:25
136.4 C 225kPa 5:55
136.3 C 224kPa 6:25
136.3 C 224kPa 6:25
VENTING 6:25
AIR DRYING 7:17
CYCLE COMPLETE 7:26

Digital Signature #
26E9090B8C6437CE
