

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3013

Name: RITTER2  
Sterilizer ID: V2313228  
10/16/2023 11:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.0 F 0.0  
2:00 183.5 F 3.5  
4:00 200.7 F 9.1  
6:00 221.6 F 7.4  
8:00 230.8 F 10.7  
10:00 233.1 F 8.6  
12:00 238.8 F 10.0  
14:00 248.7 F 15.1  
16:00 261.7 F 22.5  
18:00 272.7 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.5 F 29.6  
2:00 271.5 F 28.8  
3:00 272.5 F 29.9  
4:00 272.4 F 29.4  
  
Min 271.3 F 28.8  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.2 F 0.0  
5:00 239.2 F 0.0  
10:00 224.9 F 0.0  
15:00 227.6 F 0.0  
20:00 229.3 F 0.0  
25:00 230.4 F 0.0  
30:00 230.2 F 0.0  
Drying Complete

FILLING: 1:56  
HEATING: 18:19  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:28  
CYCLE COMPLETE

10/16/2023 12:18 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_