
Midmark M11 - v1.0.5
Total Cycles: 1591

Name: RITTER2
Sterilizer ID: V2313228
10/13/2022 07:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.5 F 0.0
2:00 179.3 F 1.9
4:00 176.0 F 5.4
6:00 199.6 F 10.0
8:00 219.7 F 6.1
10:00 236.5 F 10.1
12:00 233.4 F 8.6
14:00 244.8 F 13.2
16:00 242.9 F 12.0
18:00 255.3 F 18.5
20:00 266.3 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.5 F 29.4
2:00 270.9 F 29.0
3:00 273.1 F 29.8
4:00 272.3 F 29.9

Min 270.9 F 28.7
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.6 F 0.0
5:00 235.6 F 0.0
10:00 230.1 F 0.0
15:00 224.7 F 0.0
20:00 223.0 F 0.0
25:00 223.1 F 0.0
30:00 212.8 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 21:40
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:53
CYCLE COMPLETE

10/13/2022 08:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____