
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000282
8:33 02/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:06
STERILIZING 4:36
136.5 C 233kPa 4:36
136.1 C 230kPa 5:06
136.1 C 230kPa 5:36
136.2 C 230kPa 6:06
136.4 C 233kPa 6:36
136.3 C 232kPa 7:06
136.4 C 233kPa 7:36
136.2 C 231kPa 8:06
136.2 C 231kPa 8:06
VENTING 8:06
AIR DRYING 9:01
CYCLE COMPLETE 18:56

Digital Signature #
973B9FA2A1EB38F8
