

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3200

Name: RITTER2  
Sterilizer ID: V2313228  
12/08/2023 05:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 120.2 F 0.0  
2:00 126.0 F 1.1  
4:00 156.0 F 4.6  
6:00 182.4 F 9.3  
8:00 203.4 F 6.4  
10:00 221.3 F 7.8  
12:00 234.9 F 13.1  
14:00 237.4 F 10.5  
16:00 236.2 F 8.9  
18:00 251.1 F 16.0  
20:00 263.4 F 23.2  
22:00 271.5 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.4 F 29.0  
2:00 272.5 F 29.4  
3:00 271.8 F 28.5  
4:00 272.8 F 29.4

Min 271.5 F 28.5  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.9 F 0.0  
5:00 195.0 F 0.0  
10:00 208.8 F 0.0  
15:00 207.0 F 0.0  
20:00 207.5 F 0.0  
25:00 204.7 F 0.0  
30:00 213.0 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 22:01  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:16  
CYCLE COMPLETE

12/08/2023 06:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_