

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2575

Name: RITTER2  
Sterilizer ID: V2313228  
06/27/2023 01:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.0 F 0.0  
2:00 120.0 F 0.0  
4:00 128.2 F 0.0  
6:00 166.2 F 2.4  
8:00 196.9 F 6.1  
10:00 209.3 F 10.8  
12:00 226.4 F 5.9  
14:00 230.7 F 10.5  
16:00 233.3 F 8.4  
18:00 245.0 F 13.4  
20:00 243.5 F 12.3  
22:00 255.6 F 18.7  
24:00 266.3 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.8  
1:00 272.7 F 29.6  
2:00 272.2 F 29.2  
3:00 272.9 F 29.7  
4:00 271.9 F 29.6

Min 271.0 F 28.7  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.1 F 0.0  
5:00 214.6 F 0.0  
10:00 222.5 F 0.0  
15:00 223.6 F 0.0  
20:00 213.5 F 0.0  
25:00 219.5 F 0.0  
30:00 221.9 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 25:42  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:03:57  
CYCLE COMPLETE

06/27/2023 02:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_